




BEACON
A Beacon Group Company

BONAVERA Count 5R

5 Part Differential
Hematology Analyzer



With
RETICULOCYTE

SALIENT FEATURES

- ▶ 28 Reportable & 6 RUO* Parameters including Retic_Abs, Retic% as well as IRF
- ▶ Flow Cytometry (FCM) and Multi Angle Scatter Separation (MASS) Technology
- ▶ Unique Resistant RBC Mode
- ▶ Throughput of 60 Sample / Hr.
- ▶ Storage of 2,00,000 Results with Histogram & Scattergrams

BONAVERA Count 5R

- ▶ 28 Reportable & 6 RUO* Parameters with 2 Histograms & 2 Scattergrams
- ▶ Reticulocyte count with RETIC_ABS, Retic% & IRF
- ▶ Impedance method for RBC & PLT counting
- ▶ Combination of WIC & WOC methods for WBC Counting
- ▶ Cyanide Free Reagent for HGB measurement
- ▶ Throughput of 60 sample / Hour
- ▶ Storage of 2,00,000 results with histograms & scattergrams
- ▶ Supports Variety of External color printers
- ▶ Supports HIS/LIS & HL7 Protocols
- ▶ Compact size design





Laser Scatter

Multi – angle scatter separation technology (MASS)



Digital Sheath Flow Regulation

Use of only sheath flow reagent for WBC 5 Part
Chemical Dye free Technology



Separation System

Use of motor & syringes for separation
Ceramic Syringes & high precision motors



Electronics & Mechanical Components

Use of Best Quality Time Tested components and materials
that ensures trouble-free long-lasting performance



2D Grams

2D histograms for RBC & PLT
2D scattergrams for 5-part WBC differentials including
Eosinophil & neutrophils



Operation Interface

Simple touchscreen operations with user friendly icons
Guidance to operator through handy tool icons



Quality Control Solutions

QC programs with 4 options like L-J, X-R, X-B & X
Third Party QC solutions available

TECHNICAL SPECIFICATIONS

BONAVERA Count 5R

Parameters

WBC, LYM%, MON%, NEU%, EOS%, BAS%, LYM#, MON#, NEU#, EOS#, BAS#, RBC, HGB, HCT, MCV, MCH, MCHC, RDW_CV, RDW_SD, PLT, MPV, PDW, PCT, P_LCR, P_LCC, RETIC%, RETIC_ABS, IRF

Research Parameters : ALY#, ALY%, LIC#, LIC%, NRBC#, NRBC%

2 Histograms for RBC and PLT

2 Scattergrams : 5-part differential scattergram

Eosinophils and Neutrophil scattergram

Principles of Operation

WBC/DIFF : Flow Cytometry, Semi-conductor laser light multi-dimensional cell classification

WBC Analysis : Optical and impedance measurement

RBC/PLT Analysis : Impedance method

HGB Test : Cyanide-free reagent colorimetry

Reagent

Diluent, Sheath, Detergent, Lyse

Sampling Modes

Whole Blood : 20µL | Pre-Diluted : 20µL

Throughput

Whole Blood : Up to 60 samples / Hr.

Pre-diluted : Up to 60 samples / Hr.

Data Storage


Up to 200,000 sample results with grams

Alarms


Error Messages




Main Page



Calibration Setting



QC Setting



System Maintenance

Linearity

Parameter	Linearity Range	Acceptable Limits	Correlation Coefficient R
WBC	0x10 ⁹ /L ~ 10.0x10 ⁹ /L	±0.5x10 ⁹ /L	≥ 0.990
	10.1x10 ⁹ /L ~ 100.0x10 ⁹ /L	±5%	
RBC	0.10x10 ¹² /L ~ 100x10 ¹² /L	±0.05x10 ¹² /L	≥ 0.990
	1.01x10 ¹² /L ~ 8.00x10 ¹² /L	±5%	
HGB	0 g/L ~ 70 g/L	±2 g/L	≥ 0.990
	71 g/L ~ 250 g/L	±2%	
PLT	0x10 ⁹ /L1 ~ 100x10 ⁹ /L	±10x10 ⁹ /L	≥ 0.990
	101x10 ⁹ /L ~ 1000x10 ⁹ /L	±8%	

Display

10.4 inch with touch screen

Requirements

Voltage : 110 - 220V (AC), 1 KVA Online UPS
with Isolated Transformer

Temperature : 15 - 30

Humidity : 30 - 85%

Size : L 490 mm x H 332 mm x W 459 mm

Weight : 35 KG



FACTORY & REGISTERED OFFICE :
424, NEW GIDC, KABILPORE, NAVSARI - 396 424 (GUJ) INDIA
TEL : +91 - 74900 45631 | E-mail : exports@beaconindia.com
Website : www.beaconindia.com | CIN No. : U24232GJ1990PTC013764

